



# DECLARATION OF CONFORMITY

Apparatus Identification

**Fluke Model Fluke-312,316, & 318 Clamp Meters for China**

Apparatus Classification

**Test and Measurement Equipment**

Statement of Conformity

**Based on sample product test results using appropriate standards, and in accordance with the following EC Directives, the Fluke Corporation hereby declares the Fluke-312,316 & 318 Clamp Meters to be in conformity with:**

**EC Directive 89/336/EEC, Electromagnetic Compatibility, and  
EC Directive 73/23/EEC, Low Voltage (Amended 93/68/EEC).**

Sample Product Testing for EMC

*Tested By* **Fluke Corporation**  
PO Box 9090, Everett, WA 98206-9090

*Standards Used* **EN61326-1**, (2000) Electrical equipment for measurement, control and laboratory use.

*Report ID* For **Fluke Model Fluke-312,316 & 318 Clamp Meters for China.**

Sample Product Testing for Safety

*Tested By* **Fluke Corporation**  
P.O. Box 9090  
Everett, WA 98206-9090

*Standards Used* **EN 61010-2:032: 1994**

*Report ID* For the Fluke-312, 316, & 318 Clamp Meters for China.

Manufacturer

**FLUKE Corporation**  
6920 Seaway Boulevard  
PO Box 9090  
Everett, WA 98206-9090  
United States of America

EU Representative

**FLUKE Europe B.V.**  
Science Park Eindhoven 5110  
5692 EC SON Eindhoven  
The Netherlands

  
James Cavoretto., Vice President  
Fluke Industrial Group Division

Jan Roetman, General Manager  
Fluke Europe B.V.